


LASER BAR LST-L-3336BAS



The main defect in the work of the diode handle is that the outbreaks stop and the sensation of tingling disappears. Often the problem can be seen through the observation window.



site WWW.ROMSHI.COM

The diode stack is a frequent reason lasers stop working. Scan and visit www.romshi.com for help.