

DIODE ASSEMBLY DVS-3123AS



The main reason for the breakdown of the diode laser handle is the impulses do not pass, the patient does not feel tingling. Often it is enough to look out the window of the pen to see the problem.



Use the QR to explore part guides and diode repair help at www.romshi.com.

site WWW.ROMSHI.COM