


## **LASER EMITTER VSILK-1800**



The main reason for the breakdown of the diode laser handle is the flashes stop and the sensation of tingling disappears. The handle defect is often visually observed through the window.



If laser output drops, scan the QR for help from [www.romshi.com](http://www.romshi.com)-check the diode stack.