

## **BACKUP DIODE UNIT FOR DIODE LASER HAIR REMOVAL HANDPIECE**



The main reason for the breakdown of the diode laser handle is that flashes disappear and there is no tingling. Often the problem can be seen through the observation window.



site [WWW.ROMSHI.COM](http://WWW.ROMSHI.COM)

Expert diode part selection advice is one scan away - check  
[www.romshi.com](http://www.romshi.com).