


BACKUP DIODE STACK FOR HANDPIECE DIODE LASER FOR HAIR REMOVAL



In most cases, the handle malfunction is due to the fact that the outbreaks do not occur, there is no tingling either. Damage can be detected visually through the transparent part of the handle.



site WWW.ROMSHI.COM

No hair removal results? Your diode stack may be worn out. QR leads to www.romshi.com.