

## REPLACEMENT DIODE LASER STACK FOR HAIR REMOVAL DEVICE



The main defect in the work of the diode handle is that the impulses do not pass, the patient does not feel tingling. The handle defect is often visually observed through the window.



site [WWW.ROMSHI.COM](http://WWW.ROMSHI.COM)

Laser broken? Find out which part to replace - scan to visit  
[www.romshi.com](http://www.romshi.com).