


ADDITIONAL DIODE MODULE FOR LASER HAIR REMOVAL HAND TOOL



In most cases, the handle malfunction is due to the fact that the outbreaks do not occur, there is no tingling either. Often the problem can be seen through the observation window.



site WWW.ROMSHI.COM

Laser device issues? Diode stack may be failing. Use the QR code to access www.romshi.com for help.